



## CERTIFIED TEST REPORT

CUSTOMER: J W KERNS

MODEL NO: M0304

METER SERIAL NO: 12-07804

### CONFIGURATION

METER INSIDE DIAMETER: 3.791

METER OUTSIDE DIAMETER: 4

TEST DATE: 8/23/2012

TEST FACILITY: 5000# Gravimetric

IDEAL TEST CONSTANT: 1977

### CALIBRATION DATA

	<u>Tested TC</u>	<u>GPM</u>	<u>Accuracy</u>
1	1983	641	100.3

CERTIFIED BY: Paul Hobbs

DATE: 8/24/2012

This calibration was performed on a gravimetric or volumetric test facility, traceable to the National Institute of Standards and Technology, USA. The estimated flow measurement uncertainty of the calibration facilities are:  
Gravimetric +/- 0.15%      Volumetric +/- 0.5%



**McCROMETER**

3255 WEST STETSON AVENUE

HEMET, CA 92545 USA

PHONE (951) 652-6811 / FAX (951) 652-3078

WEB SITE: <http://www.mccrometer.com> E-MAIL: [info@mccrometer.com](mailto:info@mccrometer.com)



12-07804

8/24/2012 11:43:01 AM  
Version 1.2 (4/18/2007)