



*2nd Day
Meter*

CERTIFIED TEST REPORT

CUSTOMER: J W KERNS

MODEL NO: M0308

METER SERIAL NO: 12-07806

CONFIGURATION

METER INSIDE DIAMETER: 7.791

METER OUTSIDE DIAMETER: 8

TEST DATE: 8/23/2012

TEST FACILITY: Volumetric

IDEAL TEST CONSTANT: 3938

CALIBRATION DATA

	<u>Tested TC</u>	<u>GPM</u>	<u>Accuracy</u>
1	3962	1555	100.6

CERTIFIED BY: Paul Hobbs

DATE: 8/24/2012

This calibration was performed on a gravimetric or volumetric test facility, traceable to the National Institute of Standards and Technology, USA. The estimated flow measurement uncertainty of the calibration facilities are:
Gravimetric +/- 0.15% Volumetric +/- 0.5%



McCROMETER

3255 WEST STETSON AVENUE

HEMET, CA 92545 USA

PHONE (951) 652-6811 / FAX (951) 652-3078

WEB SITE: <http://www.mccrometer.com> E-MAIL: info@mccrometer.com



12-07806

8/24/2012 11:51:31 AM
Version 1.2 (4/18/2007)